



Declaration of Conformity

Manufacturer:

ResMed Pty. Ltd.
1 Elizabeth Macarthur Drive
Bella Vista
NSW 2153
Australia

Authorised Representative:

ResMed SAS
Parc Technologique de Lyon
292 Allée Jacques Monod
69791 Saint Priest Cedex
France

Notified Body:

TÜV SÜD Product Service
GmbH
Ridlerstraße 65
80339 München
Germany

Product: SlimLine

Intended Use: SlimLine Tubing is a non-invasive medical device accessory used for conveying the air-flow (with or without supplemental oxygen) generated by flow generators to a nasal/full face mask or nasal pillow-system for the treatment of CPAP or Bi-level therapy. It is designed to use with S9 Series, Air10 Series, Lumis and Stellar.

Classification: IIa according to Rule 2

GMDN: 37705 Breathing circuit, ventilator, reusable

Conformity Assessment Route: Annex II (excluding Section 4), 93/42/EEC

We herewith declare that the above mentioned products are in conformity with the Council Directive 93/42/EEC for medical devices including the MDD amendment 2007/47/EC.

Compliance is applicable from the date listed below. All supporting documentation is retained at the premises of the manufacturer.

This declaration is issued under the sole responsibility of ResMed Pty. Ltd.

EC Certificate Number: G1 049861 0158

Signed at Sydney, Australia on: 11 December 2019

Johanna Wright
Director of Regulatory Affairs
ResMed Pty. Ltd.

EC074

First issued: 1 April 2011